


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVL-MA3-P

S.N.: 3029

Date: APRIL 15 '86



- a) Output signal amplitude:
 $\approx +100$ VOLTS
- b) Pulse width:
 ≤ 4.0 NS
- c) Rise time:
 NA
- d) Fall time:
 NA
- e) PRF:
 0 TO 20 KHz
- f) Jitter, stability:
 OK
- g) Prime power:

100 VOLTS/DIV
2.0 NS/DIV
20 KHz

+15 \pm 1.0 VOLTS
200 MA (MAX)
[Signature]