


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVL-AV-1-C-P-W

S.N.: 3045

Date: MAY 15 66



- ) Output signal amplitude:  
0 TO 400 VOLTS
- ) Pulse width:  
3 TO 400 NSEC\*
- ) Rise time:  
≤ 1 NSEC
- ) Fall time:  
≤ 2 NSEC
- ) PRF:  
0 TO 5 KHZ
- ) Jitter, stability:  
OK
- ) Prime power:

100 VOLTS/DIV  
10 NSEC/DIV  
5 KHZ

120 / 240 V  
50 - 60 Hz

\* For 100 TO 400 NSEC  
NEEDS ADDITIONAL  
EXTERNAL DELAY LINE  
(SEE INSTRUCTIONS)