


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVP - AV - 17V3 - C - N*

S.N.: *3050*

Date: *MAY 26 86*



- a) Output signal amplitude:
0 TO -40 VOLTS.
- b) Pulse width:
0.4 TO 2.0 NS
- c) Rise time:
≤ 150 PSEC
- d) Fall time:
≤ 250 PSEC
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:

120 / 240 V

50 - 60 Hz.

*≈ 32 VOLTS/DIV
100 NS/DIV
100 kHz*

[Signature]