


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVL-2A-PS-PN-W-05-M.*  
S.N.: *3055*  
Date: *MAY 20 86*



- a) Output signal amplitude:  
*0 TO ± 170 VOLTS*
- b) Pulse width:  
*0 TO 400 NSEC\**
- c) Rise time:  
*≤ 1 NSEC*
- d) Fall time:  
*≤ 3 NSEC*
- e) PRF:  
*0 TO 5 KHZ*
- f) Jitter, stability:  
*OK.*
- g) Prime power:

*120/240 V*

*50 - 60 HZ.*

*[Signature]*

*100 VOLTS/DIV  
10 NSEC/DIV  
5 KHZ.*

*\* REQUIRES EXTERNAL  
DELAY LINE FOR  
100 TO 400 NSEC,  
SEE PARA 6.*