


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PVL-TN-1*
S.N.: *3056*
Date: *MAR 14 86*



- a) Output signal amplitude:
-250 V + 30 V OFFSET
- b) Pulse width:
5 TO 100 NSEC
- c) Rise time:
≤ 2 NSEC.
- d) Fall time:
≤ 3 NSEC.
- e) PRF:
0 TO 5 KHZ
- f) Jitter, stability:
OK
- g) Prime power:

100 VOLTS/DIV.
10 NSEC/DIV

120/240 V

50-60 Hz

h) t_p : ≤ 100 nsec.