


PULSE GENERATOR
PERFORMANCE CHECK

Model: PVI-U-1A2-PS-05-M-PN
S.N.: 3060
Date: MAY 20 '86



- a) Output signal amplitude:
0 TO ± 100 VOLTS
- b) Pulse width:
0 TO 100 NSEC
- c) Rise time:
 ≤ 0.5 NSEC
- d) Fall time:
 ≤ 2 NSEC
- e) PRF:
0 TO 5 KHz
- f) Jitter, stability:
OK
- g) Prime power:

120 / 240 V
50 - 60 Hz

100 VOLTS/DIV
5.0 NSEC/DIV
5.0 KHz

$R_L = 50$ OHMS