


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *MURD-2 PS -P*  
S.N.: *3062*  
Date: *AUG 11 '86*



- a) Output signal amplitude:  
*0 TO ±100 VOLTS*
- b) Pulse width:  
*50 NSEC TO 1.0 USEC*
- c) Rise time:  
*≤ 10 NSEC*
- d) Fall time:  
*≤ 10 NSEC*
- e) PRF:  
*0 TO 20 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*50 VOLTS/DIV*  
*200 NSEC/DIV*  
*R<sub>L</sub> = 50 Ω*  
*5.0 KHz*

*120/240 V*  
*50-60 Hz*

*[Signature]*