


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVH-S-P*

S.N.: *3065*

Date: *JUNE 10 86*



*≈ 32 VOLTS/DIV
1.0 NSEC/DIV
300 KHz*

- a) Output signal amplitude:
0 TO +15V
- b) Pulse width:
≤ 500 PSEC
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 100 MHz
- f) Jitter, stability:
OK
- g) Prime power:

*+15 ± 1.0 VOLTS
150 mA (MAX)*