


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVH-S-P*
S.N.: *3071*
Date: *JUNE 10 86*



- a) Output signal amplitude:
0 TO +15 VOLTS
- b) Pulse width:
≤ 500 PSEC.
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1 MHz
- f) Jitter, stability:
OK
- g) Prime power:

≈ 32 VOLTS / DIV
10 nSEC / DIV
500 kHz

+15 ± 1.0 VOLTS
150 mA (max)
[Signature]