


PULSE GENERATOR
PERFORMANCE CHECK


Model: *AVI-U-14U3-C-PN-M-05*

S.N.: *3075*

Date: *JUNE 27 '86*



Ⓐ *100 VOLTS/DIV*
10 NSEC/DIV
5 KHZ.



Ⓑ *AS Ⓐ BUT*
1.0 NSEC/DIV.
(RISE TIME)

- a) Output signal amplitude:
0 TO ± 100 VOLTS
- b) Pulse width:
0 TO 100 NSEC.
- c) Rise time:
≤ 200 PSEC
- d) Fall time:
≤ 3 NSEC
- e) PRF:
0 TO 5 KHZ
- f) Jitter, stability:
OK
- g) Prime power:

120/240 V
50-60
Hz

[Signature]