


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVM-1-PS-P-DW-D*


S.N.: *3085*

Date: *JUNE 17 86*



- a) Output signal amplitude:  
*0 TO ±5 VOLTS*
- b) Pulse width:  
*0.2 TO 6.0 NSEC.*
- c) Rise time:  
*≤ 100 PSEC*
- d) Fall time:  
*≤ 135 PSEC*
- e) PRF:  
*0 TO 25 KMHZ*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*5 VOLTS/DIV  
1.0 NSEC/DIV  
5.0 MHZ*

*120/240V  
50-60MHZ  
*