


PULSE GENERATOR
PERFORMANCE CHECK

Model: *ANM-1-PS-P-PA-D*

S.N.: *3086*

Date: *JUNE 17 81*



- a) Output signal amplitude:
0 TO ±5 VOLTS
- b) Pulse width:
0.2 TO 6.0 NSEC
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 133 PSEC
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:

*5 VOLTS/DIV.
1.0 NSEC/DIV
5.0 MHz*

*120/240 VOLTS
50-60 Hz*
[Signature]