


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVR-A-1-6A-PS-P

S.N.: 3110

Date: JULY 15 '86



- a) Output signal amplitude:
0 TO +200 VOLTS
- b) Pulse width:
50 NSEC TO 1.0 USEC
- c) Rise time:
(0.5% MAX DUTY CYCLE)
≤ 10 NSEC
- d) Fall time:
≤ 10 NSEC
- e) PRF:
MIN PW : 0 TO 100 KHz
MAX PW : 0 TO 5 KHz
- f) Jitter, stability:
OK
- g) Prime power:

120/240 V

50 - 60 Hz

[Signature]

50 VOLTS/DIV

200 NSEC/DIV

5 KHz

$R_L = 50 \Omega$