


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AP-AV-1-P-C*

S.N.: *3112*

Date: *JUNE 2 86*



- a) Output signal amplitude:
0 TO +10 VOLTS.
- b) Pulse width:
0.2 TO 400 NSEC
- c) Rise time:
≤ 100 PSEC.
- d) Fall time:
≤ 135 PSEC.
- e) PRF:
0 TO 1 MHZ
- f) Jitter, stability:
OK
- g) Prime power:

120/240V

50-60 HZ.

10 VOLTS/DIV
1.0 NSEC/DIV
100 KHZ.