


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVM-1-C-PW-D-M*

S.N.: *3114*

Date: *JUNE 20 66*



- a) Output signal amplitude:  
*0 TO ±5 VOLTS*
- b) Pulse width:  
*0.2 TO 6.0 USEC.*
- c) Rise time:  
*≤ 100 PSEC*
- d) Fall time:  
*≤ 135 PSEC*
- e) PRF:  
*0 TO 25 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*120/240 V*  
*50-60 Hz.*

*5 VOLTS DIV*  
*1.0 USEC/DIV*  
*14 MHz*