


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AU-107C-PS-M-451-P*


S.N.: *3123*

Date: *AUG 5 1966*



- a) Output signal amplitude:
*0 TO +10 AMPS TO
0 TO +100 VOLTS*
- b) Pulse width:
50 NSEC TO 5.0 MSEC.
- c) Rise time:
≤ 20 NSEC
- d) Fall time:
≤ 20 NSEC
- e) PRF:
0 TO 1 KHz.
- f) Jitter, stability:
OK
- g) Prime power:
*120/240 V
50-60 Hz*

ⓐ *R_L = 10 Ω
20 VOLTS/DIV
1.0 USEC/DIV
0.5 KHz*



ⓑ *R_L = 3 Ω
10 VOLTS/DIV
1.0 USEC/DIV
0.5 KHz*