


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVO-1.*  
S.N.: *3128*  
Date: *JUNE 17 88*



- a) Output signal amplitude:  
*+25 TO +100 AMPS*
- b) Pulse width:  
*25 NSEC*
- c) Rise time:  
*NA*
- d) Fall time:  
*NA*
- e) PRF:  
*0 TO 1 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*≈ 60 AMPS/DIV*  
*10 NSEC/DIV*  
*500 Hz.*

*+15 ± 1.0 VOLTS*  
*200 mA (MAX)*  
*[Signature]*