


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVO-1.*
S.N.: *3129*
Date: *JUNE 17 '86*



- a) Output signal amplitude:
+25 TO +100 AmPS
- b) Pulse width:
25 NSEC
- c) Rise time:
N/A
- d) Fall time:
N/A
- e) PRF:
0 TO 1KHz
- f) Jitter, stability:
OK
- g) Prime power:

~ 60 AmPS/DIV
10 NSEC/DIV
500 Hz

+15 ± 1.0 VOLT
200 mA (MAX)
[Signature]