


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVP-DV-1-PS-N*

S.N.: *3142*

Date: *JUNE 20 86*



- a) Output signal amplitude:
0 TO -10 VOLTS
- b) Pulse width:
0.2 TO 4.0 NSEC
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 1000/KHz
- f) Jitter, stability:
- g) Prime power:
OK

1.0 NSEC PIV
10 VOLTS/PIV
500 KHz.

120 / 240V
50-60 Hz
[Signature]