


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PK-A-C-PN*

S.N.: *3151*

Date: *JUNE 19 86*



- a) Output signal amplitude:
0 TO 115 VOLTS
- b) Pulse width:
4 TO 100 NSEC
- c) Rise time:
*(SEE FIG 1, 2)
≤ 100 PSEC*
- d) Fall time:
≤ 2 NSEC
- e) PRF:
*MIN PW: 0 TO 1 MHz
MAX PW: 0 TO 20 kHz
SEE FIG 2*
- f) Jitter, stability:
OK
- g) Prime power:

*10 VOLTS/DIV
2.0 NSEC/DIV*

*120/240 V
50-60 Hz.*

[Signature]