


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PMK-2-C-P-6A*

S.N.: *3154*

Date: *JUN 10 86*



- a) Output signal amplitude:
0 TO +20 VOLTS
- b) Pulse width:
20 TO 200 NSEC
(20% MAX DUTY CYCLE)
- c) Rise time:
≤ 3 NSEC
- d) Fall time:
≤ 3 NSEC
- e) PRF:
0 TO 10 MHz
(20% MAX DUTY CYCLE)
- f) Jitter, stability:
OK
- g) Prime power:

120/240V

50-60 Hz

[Signature]

10 VOLTS/DIV
10 NSEC/DIV
1.0 MHz