

PULSE GENERATOR  
PERFORMANCE CHECK

Model:


APR-1-PW-PS-P-EA

S.N.:

3163

Date:

MAY 22 1986



a) Output signal amplitude:

0 TO +200 VOLTS

b) Pulse width:

75 NSEC TO 5.0 USEC

c) Rise time:

$\leq 10$  NSEC

d) Fall time:

$\leq 10$  NSEC

e) PRF:

0 TO 1 KHz

f) Jitter, stability:

OK

g) Prime power:

120/240 V

50-60 Hz

50 VOLTS/DIV

1.0 USEC/DIV

500 Hz

~~Signature~~