


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVM -2-N-PS*

S.N.: *3166*

Date: *JULY 8 1966*



- a) Output signal amplitude:  
*0 TO -15 VOLTS*
- b) Pulse width:  
*0.2 TO 2.0 NSEC*
- c) Rise time:  
*≤ 100 PSEC*
- d) Fall time:  
*≤ 135 PSEC*
- e) PRF:  
*0 TO 25 MM Hz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*1.0 NSEC/DIV*  
*10 VOLTS/DIV*  
*5.0 MM Hz*

*120/240 V*  
*50-60 Hz*  
*[Signature]*