


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PR-A-2-PW-C-PN*

S.N.: *3198*

Date: *JULY 30 86*



- a) Output signal amplitude:
0 TO ± 50 VOLTS
- b) Pulse width:
20 TO 200. NSEC
- c) Rise time:
*(5% MAX DUTY CYCLE)
≤ 5 NSEC*
- d) Fall time:
≤ 5 NSEC
- e) PRF:
*0 TO 10 MHz.
(5% MAX DUTY CYCLE)*
- f) Jitter, stability:
OK
- g) Prime power:

*10 VOLTS/DIV
50 NSEC/DIV
200 KHz.*

*120/240 V
50-60 Hz*

[Signature]