


PULSE GENERATOR  
PERFORMANCE CHECK


Model: *ACORN 1-N*

S.N.: *3204*

Date: *Aug 7 '86*



Ⓐ *5 VOLTS/Div*  
*10 nSEC/Div*  
*≈ 50 MHz*



Ⓑ *5 VOLTS/Div*  
*1.0 nSEC/Div*  
*≈ 5 MHz*

- a) Output signal amplitude:  
*0 TO -5 VOLTS*
- b) Pulse width:  
*0 TO 6 nSEC*
- c) Rise time:  
*≤ 100 pSEC*
- d) Fall time:  
*≤ 200 pSEC*
- e) PRF:  
*0 TO 50 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*+24 ± 1.0 VOLTS*  
*400 mA (MAX)*

*[Signature]*