


PULSE GENERATOR  
PERFORMANCE CHECK


Model: *AVI-V-MU1-0-P*

S.N.: *3205*

Date: *JULY 30 86*



Ⓐ *32 V/DIV*  
*5.0 NSEC/DIV*



Ⓑ *AS Ⓐ BUT*  
*1.0 NSEC/DIV*  
*(RISE TIME)*

- a) Output signal amplitude:  
*0 TO +50 VONB*
- b) Pulse width:  
*2 TO 100 NSEC*
- c) Rise time:  
*≤ 250 PSEC.*
- d) Fall time:  
*≤ 2 NSEC*
- e) PRF:  
*0 TO 20 KHz.*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240 V*  
*50-60 Hz.*

*[Signature]*