


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVP-AV-1-PS-N*

S.N.: *3217*

Date: *JUNE 20 86*



- a) Output signal amplitude:
0 TO -10 VOLTS
- b) Pulse width:
0.2 TO 4.0 NS
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:

120/240V

50-60 Hz

[Signature]

*10 VOLTS/DIV
1.0 NS/DIV
500 kHz*