


PULSE GENERATOR
PERFORMANCE CHECK

Model: *MP-M-143 P*

S.N.: *3223*

Date: *AUG 13 86*



- a) Output signal amplitude:
0 TO +40 VOLTS
- b) Pulse width:
0.4 TO 2.0 NS
- c) Rise time:
≤ 150 NS
- d) Fall time:
≤ 250 NS
- e) PRF:
0 TO 1 MHz
- f) Jitter, stability:
OK
- g) Prime power:

+15 ± 1.0 VOLT
200 mA (MAX)

2.32 VOLTS/DIV
1.0 NS/DIV
500 KHz