


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *Av14-MPT-T*

S.N.: *3224*

Date: *Aug 13 '86*



- a) Output signal amplitude:  
*≈ +30 VOLTS*
- b) Pulse width:  
*≈ 1 NSEC.*
- c) Rise time:  
*NA*
- d) Fall time:  
*NA*
- e) PRF:  
~~*1 MHz*~~
- f) Jitter, stability:  
*OK*
- g) Prime power:

*32 VOLTS PIV*  
*100 NSEC PIV*  
*500 KHZ*

*+15 ± 1.0 VOLT*  
*150 mA.*