


PULSE GENERATOR
PERFORMANCE CHECK


Model: AVRL-1TT8B

S.N.: 3230

Date: SEPT 2 86



A OUT
50 VOLTS/DIV
100 NSER/DIV



B OUT
50 VOLTS/DIV
100 NSER/DIV

- a) Output signal amplitude:
-200 VOLTS TO
- b) Pulse width: 1K OR 144.
A) 10 TO 100 NSER
B) 10 TO 100 NSER
- c) Rise time:
≤ 5 NSER
- d) Fall time:
≤ 5 NSER
- e) PRF: 60 Hz
- f) Jitter, stability:
OK
- g) Prime power:

120/240 V
50-60 Hz

- h) DELAYS:
A) 200 NSER TO 2.2 μs
B) 16.6 μSER + 0.2 μSER
TO 16.6 μSER + 2.2 μs

[Signature]