



PULSE GENERATOR
PERFORMANCE CHECK

Model: AVR-ITT 813
S.N.: 3232
Date: SEPT 30 86



A PULSE 50 VOLTS/DIV
100 NS/2 DIV



B PULSE 50 VOLTS/DIV
100 NS/2 DIV

- a) Output signal amplitude: ~200 VOLTS
- b) Pulse width: 10 TO 100 NS
- c) Rise time: 5 NS
- d) Fall time: 5 NS
- e) PRF: 60 Hz
- f) Jitter, stability: OK
- g) Prime power: 120 / 240 V
50 - 60 Hz