


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVMH-3-C-PN

S.N.: 3237

Date: APRIL 22 87



50 dB ATTEN

32 VOLTS/DIV

10 MHz

- a) Output signal amplitude:
0 TO ± 50 VOLTS
- b) Pulse width:
≤ 600 PSEC
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:

120/240V

50-60 Hz

[Signature]