


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AMP-MV-143-P*

S.N.: *3255*

Date: *SEPT 30 86*



*= 32 VOLTS/DIV
1.0 NSEC/DIV
100 KHz*

- a) Output signal amplitude:
0 TO +40 VOLTS
- b) Pulse width:
0.4 TO 2.0 NSEC
- c) Rise time:
≤ 150 NSEC
- d) Fall time:
≤ 250 NSEC
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
*+15 ± 1.0 VOLT
250 mA (MAX)*
[Signature]