


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVMAN-3A-C-P-1N*  
S.N.: *3267*  
Date: *OCT 10 86*



- a) Output signal amplitude:  
*0 TO 15 VOLTS*
- b) Pulse width:  
*5 TO 40 NS*
- c) Rise time:  
*6 NS*
- d) Fall time:  
*5 NS*
- e) PRF:  
*0 TO 50 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*5 VOLTS/DIV*  
*10 NS/DIV*

*120/240 V*  
*50-60 Hz*

*[Signature]*