


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *PKR-1-PW-C-P*  
S.N.: *3270*  
Date: *OCT 7 '86*



- a) Output signal amplitude:  
*0 TO +200 V.*
- b) Pulse width:  
*0.1 TO 50  $\mu$ SEC*
- c) Rise time:  
 *$\leq 10$  NSEC*
- d) Fall time:  
 *$\leq 10$  NSEC*
- e) PRF:  
*0 TO 1 KHZ.*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*50 VOLTS/DIV*  
*1.0  $\mu$ SEC/DIV*  
 *$R_L = 50 \Omega$*

*120/280V*  
*50-60 HZ.*

*[Signature]*