


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AMP-M-2-C-CA-W-PN-OS-M*

S.N.: *3276*

Date: *OCT 6 86*



*10. VOLTS/PW*  
*5 NS/PW*  
*50 KHZ*

a) Output signal amplitude:

*0 TO ±10 VOLTS*

b) Pulse width:

*20 TO 500 NS*

c) Rise time:

*≤ 100 PSEC*

d) Fall time:

*≤ 135 PSEC*

e) PRF:

*0 TO 50 KHZ*

f) Jitter, stability:

*OK*

g) Prime power:

*120 / 240 V*

*50 - 60 HZ*