


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AMH-1A23-C-P*

S.N.: *3340*

Date: *AUG 29 86*



- a) Output signal amplitude:  
*0 TO +60 VOLTS*
- b) Pulse width:  
*≤ 1.0 NSEC*
- c) Rise time:  
*NA*
- d) Fall time:  
*NA*
- e) PRF:  
*0 TO 1 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*≈ 32 VOLTS/DIV  
1.0 NSEC/DIV  
100 KHz*

*120/240 V  
50-60 Hz*  
