


PULSE GENERATOR
PERFORMANCE CHECK

Model: *ANM-1-C-N-PN-ET-EW-ED-EO*

S.N.: *3341*

Date: *OCT 15 '86*



- a) Output signal amplitude:
0 TO ± 5 VOLTS
- b) Pulse width:
0.2 TO 60 NSOR.
- c) Rise time:
≤ 100 PSOR.
- d) Fall time:
≤ 135 PSOR.
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:

5 VOLTS/DIV
1.0 NSOR/DIV
12 MHz

120/240 V
50-60 Hz
[Signature]