


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AMK-33-C-W-PN*


S.N.: *3350*

Date: *OCT 31 86*



P_{OUT} 20 VOLTS/DIV
1.0 USEC/DIV
R_L = 50 Ω
5.0 KHz

- a) Output signal amplitude:
0 TO ±100 VOLTS
- b) Pulse width:
0.1 TO 5.0 USEC.
- c) Rise time:
≤ 5 NSEC
- d) Fall time:
≤ 5 NSEC.
- e) PRF:
0 TO 5 KHz
- f) Jitter, stability:
OK
- g) Prime power:



V_{OUT} 20 VOLTS/DIV
100 NSEC/DIV
R_L = 50 Ω
5.0 KHz

120/240V
50-60Hz
[Signature]