


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVL-275-N*

S.N.: *3355*

Date: *DEC 12 86*



*70 dB ATTEN  
2 320 VOLTS/DIV  
2 KHz*

- a) Output signal amplitude:  
*0 TO -350 VOLTS*
- b) Pulse width:  
*5 TO 100 NSEC.*
- c) Rise time:  
*≤ 15 NSEC*
- d) Fall time:  
*≤ 3 NSEC*
- e) PRF:  
*0 TO 5 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*120/240 V  
50-60 Hz*

*[Signature]*