


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVO-5-PS-N*  
S.N.: *3362*  
Date: *NOV 18 1986*



- a) Output signal amplitude:  
*0 TO -28 AMPS*
- b) Pulse width:  
*6 TO 100 NSEC*
- c) Rise time:  
*≤ 3 NSEC*
- d) Fall time:  
*≤ 4 NSEC*
- e) PRF:  
*0 TO 5 KHZ*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*10 NSEC/DIV*  
*100 VOLTS/DIV*  
*R<sub>in</sub> = 3 Ω*  
*(≈ 33 AMPS/DIV)*  
*2 KHZ*

*120/280V*  
*50-60 KHZ*  
*[Signature]*