


PULSE GENERATOR

PERFORMANCE CHECK

Model: *ACM-2.C-P-PN-M-D*

S.N.: *3363*

Date: *SEPT 16 86*



- a) Output signal amplitude:
0 TO ± 15 VOLTS
- b) Pulse width:
0.2 TO 2.0 NSEC.
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 25 MHz.
- f) Jitter, stability:
OK
- g) Prime power:

120 / 240 V
50 - 60 Hz.

10 VOLTS/DIV
1.0 NSEC/DIV
≈ 12 MHz.

[Signature]