


PULSE GENERATOR  
PERFORMANCE CHECK


Model: *AVM-3-C-N-PN-OT-M*

S.N.: *3366*

Date: *NOV 24 86*



Ⓐ 40 db ATTEN  
12 AMPER



Ⓑ 40 db ATTEN  
12 AMPER

- a) Output signal amplitude:  
*0 TO ±15 VOLTS*
- b) Pulse width:  
*2 TO 15 NSEC*
- c) Rise time:  
*≤ 150 PSEC*
- d) Fall time:  
*≤ 600 PSEC*
- e) PRF:  
*0 TO 25 AMPER*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120 / 240 V*  
*50 - 60 MHz*
- h) OFFSET:  
*0 to ±5 VOLTS*