

PULSE GENERATOR
PERFORMANCE CHECK

Model:


AVO 1 - GED1

S.N.:

3368

Date:

OCT 8 86



a) Output signal amplitude:

+25 TO +100 AMPS

b) Pulse width:

15 NSEC

c) Rise time:

NA

d) Fall time:

NA

e) PRF:

0 TO 1KHZ

f) Jitter, stability:

OK

g) Prime power:

+15 ± 1.0 VOLT

250 mA (MAX)

10. NSEC/DIV

260 AMPS/DIV

500 HZ