


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AV14-C-P*  
S.N.: *3370*  
Date: *NOV 20 '86*



*1.0 nsec/div*  
*16 volts/div*  
*(50 db atten)*

- a) Output signal amplitude:  
*0 to +30 0.0475*
- b) Pulse width:  
*5 nsec*
- c) Rise time:  
*NA*
- d) Fall time:  
*NA*
- e) PRF:  
*0 to 1.0 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*120/280 V*  
*50-60 Hz*

*[Signature]*