

PULSE GENERATOR
PERFORMANCE CHECK

Model:


AVH-MA2-P

S.N.:

3371

Date:

SEPT 23 86



a) Output signal amplitude:

71 ± 60 VOLTS

b) Pulse width:

≤ 2 NSEC

c) Rise time:

NA

d) Fall time:

NA

e) PRF:

0 TO 1 MHz

f) Jitter, stability:

OK

g) Prime power:

435 ± 5 VOLTS

5 mA (MAX)

32 VOLTS/DIV

2.0 NSEC/DIV

500 kHz