


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVIT M2-P
S.N.: 3372
Date: SEPT 23 '86



- a) Output signal amplitude: $> +60$ VOLTS
- b) Pulse width: ≤ 2 NSEC
- c) Rise time: NA
- d) Fall time: NA
- e) PRF: 0 TO 1 MHz
- f) Jitter, stability: OK
- g) Prime power:

32 VOLTS/DIV
2.0 NSEC/DIV
500 KHZ

$\pm 135 \pm 5$ VOLTS
5 mA (MAX)