


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVH-102-P*

S.N.: *3374*

Date: *SEPT 23 '86*



- b) Output signal amplitude:
≥ +60 VOLTS
- c) Pulse width:
≤ 2 NSEC
- d) Rise time:
N/A
- e) Fall time:
N/A
- f) PRF:
0 TO 1 MHz
- g) Jitter, stability:
OK
- h) Prime power:

32 VOLTS PIV
2.0 NSEC
500 KHz

+135 ± 5 VOLTS
5 mA (MAX)