


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AV4-M2-P*

S.N.: *3375*

Date: *SEPT 23 '86*



- a) Output signal amplitude:
≥ +60 VOLTS
- b) Pulse width:
≤ 2 nSEC
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 100KHz
- f) Jitter, stability:
OK
- g) Prime power:

*32 VOLTS/DIV
2.0 nSEC/DIV
500 KHz*

*+135 ± 5 VOLTS
5 mA (MAX)
*[Signature]**