


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVO-1-P*
S.N.: *3376*
Date: *AUG 25 86*



- a) Output signal amplitude:
+25 TO +100 VPPS
- b) Pulse width:
≤ 25 NSEC.
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1 KHz
- f) Jitter, stability:
OK
- g) Prime power:

≈ 60 AMPs PIV
10 NSEC/PIV
1.0 KHz

+15 ± 1.0 VOLT
200 mA (max)